

Ambulatory Clinic Policy and Procedure

Title: DECONTAMINATION OF REUSABLE INSTRUMENTS	Policy Number: EP 7.20
Regulation: Joint Commission, CDC, ANSI/AAMI ST79:2010. Reference:	Effective Date: 12/2020

Policy Statement:

It is the policy of the Texas Tech University Health Sciences Center El Paso (TTUHSC El Paso) to provide evidence based guidelines for staff performing pre-cleaning, cleaning and decontamination of instruments and to identify employee safety practices.

Scope and Distribution:

This policy applies and will be distributed to all TTUHSC El Paso Ambulatory Clinics.

Definitions:

SPAULDING CLASSIFICATION: The classification system divides medical devices into categories based on the risk of infection involved with their use. This classification system is widely accepted and is used by the Food and Drug Administration (FDA), the Centers for Disease Control and Prevention (CDC), epidemiologists, microbiologists, and professional medical organizations to help determine the degree of disinfection or sterilization required for various medical devices. Three categories of medical devices and their associated level of disinfection are recognized. Critical, semi-critical and Non-critical.

CRITICAL: Instruments or devices introduced directly into the human body or into contact with the bloodstream or normally sterile areas. Such devices should be sterilized (e.g. surgical instruments, biopsy forceps).

SEMI-CRITICAL: Instruments or devices that come into contact with intact skin or mucous membranes and do not ordinarily penetrate sterile tissue. These devices should receive at least High-level disinfection, (e.g. respiratory equipment, surgical mirror, flexible endoscopes).

NON-CRITICAL: Instruments or devices that do not ordinarily touch the patient or touch only intact skin. These devices should be cleaned by disinfection. (E.g. bedrails, blood pressure cuffs).

PRE-CLEANING: procedure done to remove and loosen debris before manual/automated cleaning is performed.

CLEANING: Is the removal of visible soil (e.g. organic and inorganic material) from objects and surfaces and normally is accomplished manually or mechanically using water with detergents or enzymatic products.

DECONTAMINATION: Removes pathogenic microorganisms from objects so they are safe to handle (i.e., safe in the context of being reasonably free from a risk of disease transmission).

DISINFECTION: Describes a process that eliminates many or all pathogenic microorganisms, except bacterial spores, on inanimate objects.

HIGH-LEVEL DISINFECTION (HLD): destruction of all vegetative microorganisms, mycobacterium, small or non-lipid viruses, medium or lipid viruses, fungal spores, and some bacterial spores. Most high-level disinfectants have the ability to sterilize given sufficient exposure time.

Ambulatory Clinic Policy and Procedure

STERILIZATION: Describes a process that destroys or eliminates all forms of microbial life and is carried out in health-care facilities by physical or chemical methods. There are several types of sterilization processes available, including steam under pressure, ozone, ethylene oxide gas, hydrogen peroxide gas plasma (e.g. Sterrad).

The decision on the reprocessing method must be made in accordance with material compatibility, manufacturer's recommendations and national standard and guidelines. For thermostable equipment, sterilization should be preferred to disinfection.

Procedure:

1. Assigned clinic employees are responsible for cleaning, reprocessing, and preparing reusable devices for patient care. The process should be consistent throughout the campus. This includes standardized reprocessing steps for cleaning, decontamination, and/or sterilization. Written policy and procedure should be established and reviewed regularly.
2. Assigned employees will be trained in these practices during orientation. Ongoing educational programs and annual competencies will be implemented to foster a safe environment for patients and healthcare employees.
3. All clinic employees responsible for cleaning and reprocessing of instruments will follow the guidelines on Attachment A of this policy.
4. Detailed step by step instructions for instrument reprocessing should be displayed/available in the processing area. These instructions and any revision to them should be approved by the Infection Control Nurse.
5. Safety considerations - Only individuals properly trained in cleaning, decontamination, and handling of chemical solutions should process instruments. Work in a well-ventilated area to prevent formation of vapor. Users should refer to Safety Data Sheet (SDS) for first aid measures.

References:

- 1- Rutala WA, Weber DJ, and the Healthcare Infection Control Practices Advisory Committee. CDC guideline for disinfection and sterilization in healthcare facilities, 2008. Available at: https://www.cdc.gov/hicpac/pdf/guidelines/Disinfection_Nov_2008.pdf (Accessed Nov/Dec 2016).
- 2- 2012 Association for the Advancement of Medical Instrumentation. ANSI/AAMI ST79:2010 & A1:2010 & A2:2011 & A3:2012.
- 3- Steris products Literature, Products Label and MSDS. Available at: <https://www.steris.com/products/cleaners-and-disinfectants> (Accessed Nov/Dec 2016).
- 4- FDA Web site: <http://www.fda.gov/medicaldevices/deviceregulationandguidance/reprocessingofsingle-usedevices/ucm133514.htm> (Accessed Dec 2016).

Ambulatory Clinic Policy and Procedure

Policy Number: EP 7.20	Original Approval Date: 05/2015
Version Number: 6	Revision Date: 12/2020
Signatory approval on file by:	<p>Juan Figueroa, M.D. Director of Clinical Operations Clinic Operations Committee, Chair Texas Tech University Health Sciences Center El Paso</p> <hr/> <p>Armando Meza M.D Infection Control Committee Chairman Texas Tech University Health Sciences Center El Paso</p>